

# METRAOHM 10

# Low-Impedance Measuring Instrument

3-447-117-03 1/2.24

- Low-impedance measuring instrument for protective conductors and equipotential bonding conductors
- 200 mA measurement via battery supply
- 10 A measurement via mains supply
- Graphic colour display
- 4 mm safety measuring sockets
- Integrated memory
- USB port
- Software tool for reading out measured values



## **Applications**

METRAOHM 10 is a low-impedance measuring instrument for protective conductors and equipotential bonding conductors. It is designed to perform two different measurement types:

- R<sub>LO</sub> four-wire measurement with current 10 A AC
- R<sub>LO</sub> two-wire measurement with current 200 mA DC
- Contact resistances on lightning protection and grounding systems in accordance with DIN 18014
- Low-resistance continuity on machines in accordance with EN 60204-1 (VDE 0113-1)

#### **Features**

- Easy to handle device
- Simple operation with pushbuttons
- Rugged plastic housing
- 200 mA measurement in auto mode or continuous mode
- 10 A measurement in auto mode or manual mode
- Measuring resistance for checking the Kelvin clips and determining the offset values
- Data export to PC (Microsoft Excel-files)
- Data evaluation via software

# **METRAOHM 10**

Low-Impedance Measuring Instrument

# **Technical Data**

### **Mechanical Design**

#### **Relevant Standards**

Mechanical Design		The instrument has been manufactured and tested in accordance with the following safety regulations:	
Housing (W × H × D) Weight	approx. 25 cm × 10.7 cm × 13.5 cm approx. 1.75 kg (with batteries, without cables)	DIN EN 61010-1 IEC 61010-1	Safety requirements for electrical equipment for measurement, con- trol and laboratory use – Part 1: General requirements
Display Ambient Condition	TFT LCD display, 480 × 320, 3.5" S	IEC/EN 61557-1	Electrical safety in low voltage distribution systems up to 1000 $V_{AC}$ and 1500 $V_{DC}$ – Equipment for
Operating tempera- ture	+0 +40 °C		testing, measuring or monitoring of protective measures – Part 1: General requirements
Storage temperature Relative humidity	<ul> <li>-10 +70 °C (without batteries)</li> <li>max. 80%,</li> <li>no condensation allowed</li> </ul>	IEC/EN 61557-4	Electrical safety in low voltage distribution systems up to 1000 $V_{AC}$ and 1500 $V_{DC}$ – Equipment for
Elevation	max. 2000 m		testing, measuring or monitoring of
Place of use	Indoor use only		protective measures – Part 4: Resistance of earth con- nection and equipotential bonding
Power Supply		DIN EN IEC 61326-1	Electrical equipment for measure-
Battery	6 × LR6 AA alkaline battery Voltage: 1.5 V		ment, control and laboratory use - EMC requirements –
Mains power supply	230 V <sub>AC</sub> 50 Hz		Part 1: General requirements
Electrical Safety		DIN EN 60529 IEC 60529	Test instruments and test proce- dures
Measuring category	Cat II / 600 V		Degrees of protection provided by enclosures (IP code)

#### Elec

Measuring category	Cat II / 600 V
Pollution degree	2
Protection category	Cat II
Fuses	F1 (10 A measurement): 10 A / 600 V, 6.3 mm × 32 mm F2 (200 mA measurement): 400 mA / 600 V, 6.3 mm × 32 mm F3 (mains): 315 mA / 250 V, 5 mm × 20 mm

#### **Data Interfaces**

COM Port USB Type B Connection

## **Features Included**

- \_  $R_{LO}$  four-wire measurement with current 10  $A_{AC}$
- R<sub>LO</sub> two-wire measurement with current 200 mA<sub>DC</sub>

## **Characteristic Values**

R <sub>LO</sub> 200 mA (Auto mode)		
Measuring range	0.01 9.99 Ω	
	10.0 99.9 Ω	
Resolution	0.01 Ω	
	0.1 Ω	
Accuracy	± (2 % m.v. + 2 dgt)	
Test current	> 200 mA <sub>DC</sub> for $R \le 5 \Omega$ (including compensation value)	
Test current resolution	1 mA	
Open circuit voltage	$4 \text{ V} \le \text{V}_0 \le 12 \text{ V}_{\text{DC}}$	
Measurement method	2 wires	
Standard	IEC/EN 61557-4 (for R > 0.10 Ω)	

# **METRAOHM 10**

# Low-Impedance Measuring Instrument

R <sub>LO</sub> 200 mA (Continuous mode)		
Measuring range	0.01 9.99 Ω	
	10.0 99.9 Ω	
Resolution	0.01 Ω	
	0.1 Ω	
Accuracy	± (2 % m.v. + 2 dgt)	
Test current	> 200 mA <sub>DC</sub> for R $\leq$ 5 $\Omega$ (including compensation value)	
Test current resolution	1 mA	
Open circuit voltage	$4 \text{ V} \le \text{V}_0 \le 12 \text{ V}_{DC}$	
Measurement method	2 wires	
Continuous mode	Cont+ (R+) and Cont- (R-)	
Duration of continuous test	1 15 s	
Standard	IEC/EN 61557-4 (for R > 0.10 Ω)	

R <sub>LO</sub> 10 A Limit MAN		
Measuring range	0.001 0.999 Ω	
Resolution	0.001 Ω	
Accuracy	± (1 % m.v. + 2 dgt)	
Test current	$>$ 10 $A_{AC}$ for R $\leq$ 0.45 $\Omega$	
Test current resolution	0.1 A	
Open circuit voltage	$V_0 \le 12 V_{AC}$	
Measurement method	4 wires	
Power supply	230 V AC 50 Hz	
Standard	IEC/EN 61439-1	

R <sub>LO</sub> 10 A Limit AUTO		
Measuring range	0.001 0.999 Ω	
Resolution	0.001 Ω	
Accuracy	± (1 % m.v. + 2 dgt)	
Test current	> 10 A <sub>AC</sub> for R $\leq$ 0.45 $\Omega$	
Test current resolution	0.1 A	

 $\label{eq:section} \begin{array}{|c|c|} \hline \text{Open circuit voltage} & V_0 \leq 12 \ \text{V}_{AC} \\ \hline \text{Measurement method} & 4 \ \text{wires} \\ \hline \text{Selectable section} & 0.5, 1.0, 1.5, 2.5, 4.0, 6.0, 10 \\ \text{and 16 } \text{mm}^2 \\ (p[\text{Cu}] = 0.017 \ \Omega \ \text{mm}^2/\text{m}) \\ \hline \text{Length measuring} \\ \hline \text{range} & 0.1 \ \text{m} \ \dots \ 999.9 \ \text{m} \\ \hline \text{Power supply} & 230 \ \text{V}_{AC} \ 50 \ \text{Hz} \\ \hline \text{Standard} & \text{IEC/EN } 60204\text{-1} \\ \hline \end{array}$ 

### **Scope of Delivery**

1

1

1

1

1

1 1

- 1 METRAOHM 10 (M630K)
  - AC mains power supply cable
  - USB type A to type B cable
  - Two-wire measurement standard test leads with alligator clips
- 1 Four-wire Kelvin test leads with alligator clips
  - Kelvin clip test 1 Ω / 25 W (Z630K)
  - Carrying case with carrying strap and strap on the instrument
  - Test report
  - Calibration certificate

### **Order Information**

Description		Article number
METRAOHM 10	Low-Impedance Measuring Instrument	M630K

For additional information regarding accessories please refer to

- Measuring Instruments and Testers catalog
- www.gossenmetrawatt.com

© Gossen Metrawatt GmbH

Prepared in Germany • Subject to change, errors excepted • PDF version available on the Internet

All trademarks, registered trademarks, logos, product names, and company names are the property of their respective owners.

Gossen Metrawatt GmbH Südwestpark 15 90449 Nürnberg Germany Telefon +49 911 8602-0 Telefax +49 911 8602-669 E-Mail info@gossenmetrawatt.com www.gossenmetrawatt.com